

7th World Congress on

Mass Spectrometry

June 20-22, 2018 | Rome, Italy

Development of on-line and field dual TD GC – MS for automatic VOC monitoring on petrochemical sites

Damien Bazin, Michel Robert and Franck Amiet
Chromatotec, France

Ambient air is polluted by many semi-volatile and volatile organic compounds (SVOCs and VOCs) coming from anthropogenic and natural sources. VOCs from PAMS, TO14 and TO15 lists can be measured in many ranges of concentration, from pg/m³ up to mg/m³ depending on the location of the measurements. Due to the large number of molecules, the complete separation of all compounds is difficult to perform using chromatographic columns. Therefore, it is difficult to quantify precisely all compounds using a chromatograph equipped with a nonspecific detector. The goal of this study is to perform automatic and continuous identification and quantification of SVOCs and VOCs using a dual thermal-desorber gas chromatograph equipped with two flame ionization detectors (FIDs) and one mass spectrometer (MS). The device is required to identify automatically co-eluted compounds by MS technology adapted to industrial context. The coupling of two different FID GCs to a single Quadrupole MS allowed by an elaborated multiplexing system is the originality of the project: one thermal desorption gas chromatography (TDGC) for light compounds (C₂-C₅) and one TDGC for heavy compounds (C₆-C₁₆) with specific analytical conditions for each system. A measurement campaign in petrochemical new site shows the concentration of about 100 compounds at different steps of the commissioning. Thanks to the continuous measurements every 30 minutes, variations of concentrations during specific hours can be monitored. Potentially co-eluted compounds like terpenes or organochlorinated are identified and monitored at ppt level. This fully automatic system allows non-specialist operators to access expertise level results.

Biography

Damien Bazin received his PhD in Physical Chemistry at the University of Bordeaux I. He worked at the CNRS under the supervision of Pr Chrystel Faure and published major publications on the preparation and characterization of structured surfaces. He is now a Product Manager at Chromatotec® and works on the development of GC-MS/FID instruments and applications.